

Certificate of Maintenance

Certificate Number: DLCC202510184269

Dustlight name	Dustlight-AFX	Maintenance facility	airfilter.expert
Old sensor ID	2C5F8C172DB3397B	New sensor ID	D37D10C31D012D4B
Old firmware version	2.14.0	New firmware version	2.14.0
Serial number	DL01-0248351474	Maintenance date	18.10.2025
MAC address	48:27:E2:EO:C9:14	Next maintenance	18.10.2026

This maintenance was performed by: Alexander Mai

Scope of work:

- Visual inspection and cleaning of the device
- Replacement of the sensor unit with a new, factory-calibrated unit
- Replacement or cleaning of the pre-filters as required
- Functional testing of all hardware components
- Firmware update (if required)

Calibration:

The sensor unit of the Dustlight was replaced by a recalibrated unit as part of the maintenance. This goes through the same calibration process as a newly purchased Dustlight. This means that all measurement deviations are now within the permissible range according to the specification.



This Certificate of Maintenance confirms the proper testing, maintenance, and calibration of the aforementioned Dustlight. All steps were carried out according to Dustlight's specifications in a licensed workshop.

Next maintenance interval: **18.10.2026**